

UNITED STATES PATENT AND TRADEMARK OFFICE



UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/724,821	12/02/2003	Dan Bourla	1371DAN-US	2742
32964 7590 07/10/2007 DEKEL PATENT LTD., DAVID KLEIN			EXAMINER	
BEIT HAROF'IM			THOMAS, BRANDI N	
18 MENUHA VENAHALA STREET, ROOM 27 REHOVOT, 76209		ART UNIT	PAPER NUMBER	
ISRAEL	5207	•	2873	
			MAIL DATE	DELIVERY MODE
			07/10/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	Application No.	Applicant(s)				
	10/724,821	BOURLA, DAN				
Office Action Summary	Examiner	Art Unit				
	Brandi N. Thomas	2873				
The MAILING DATE of this communication appears on the cover sheet with the correspondence address						
Period for Reply A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS,						
WHICHEVER IS LONGER, FROM THE MAILING DA - Extensions of time may be available under the provisions of 37 CFR 1.13 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period w - Failure to reply within the set or extended period for reply will, by statute, Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be tim vill apply and will expire SIX (6) MONTHS from cause the application to become ABANDONEI	. the mailing date of this communication. (35 U.S.C. § 133).				
Status						
1) Responsive to communication(s) filed on 02 De	Responsive to communication(s) filed on <u>02 December 2003</u> .					
,	,—					
,—	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is					
closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.						
Disposition of Claims						
4) Claim(s) 7-16 is/are pending in the application.						
4a) Of the above claim(s) is/are withdrawn from consideration.						
5) Claim(s) is/are allowed.						
_	6) Claim(s) 7-16 is/are rejected.					
7) Claim(s) is/are objected to. 8) Claim(s) are subject to restriction and/or election requirement.						
ordinities) are subject to restriction and/or election requirement.						
Application Papers						
9) The specification is objected to by the Examiner.						
10)⊠ The drawing(s) filed on <u>02 December 2003</u> is/are: a)⊠ accepted or b)□ objected to by the Examiner.						
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).						
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
		,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,				
Priority under 35 U.S.C. § 119						
12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).						
a) All b) Some * c) None of:						
 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No 						
3. Copies of the certified copies of the priority documents have been received in this National Stage						
application from the International Bureau (PCT Rule 17.2(a)).						
* See the attached detailed Office action for a list of the certified copies not received.						
Attachment(s)						
1) Notice of References Cited (PTO-892) 4) Interview Summary (PTO-413)						
 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO/SB/08) 	Paper No(s)/Mail Da 5) Notice of Informal F	atent Application				
Paper No(s)/Mail Date 6) Other: <u>Detailed Action</u> .						

Application/Control Number: 10/724,821

Art Unit: 2873

DETAILED ACTION


Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

- (b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.
- 2. Claims 7-10 are rejected under 35 U.S.C. 102(b) as being anticipated by Fercher (5847827).

Regarding claim 7, Fercher discloses, in figures 1 and 7, a method for performing biometry, comprising: removing a lens from an eye; and making biometric measurements with a partial coherence interferometry (PCI) device aimed at the eye (30) after removal of the lens therefrom (col.9, lines 51-59).



Regarding claim 8, Fercher discloses, in figures 1 and 7, a method for performing biometry, further comprising calculating optical features based on the biometric measurements

Art Unit: 2873

and selecting an intraocular lens (12) (IOL) in accordance with the optical features (col. 10, lines 5-10).

Regarding claim 9, Fercher discloses, in figures 1 and 7, a method for performing biometry, further comprising making biometric measurements with the PCI device aimed at the eye (30) after insertion of an IOL (12) into the eye (30) (col. 10, lines 5-10).

Regarding claim 10, Fercher discloses, in figures 1 and 7, a method for performing biometry, further comprising making pre-incision biometric measurements with the PCI device (col. 9, lines 50-54).


Claim Rejections - 35 USC § 103

- 3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
 - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.
- 4. Claims 11-16 are rejected under 35 U.S.C. 103(a) as being unpatentable over Fercher (5847827) as applied to claim 7 above, and further in view of Baumann et al. (2006/0146283).

Regarding claim 11, Fercher discloses, in figures 1 and 7, a method for performing biometry but does not specifically disclose using said microscope to focus radiation from the PCI device to the eye. Baumann et al. discloses using said microscope (2) to focus radiation from the PCI device to the eye (12) (section 0017). Therefore it would have been obvious to one having ordinary skill in the art at the time the invention was made to combine the device of Fercher with the microscope of Baumann et al. for the purpose of using the microscope as an observation system (section 0017).

Application/Control Number: 10/724,821

Art Unit: 2873



Regarding claim 12, Fercher discloses, in figures 1 and 7, a method for performing biometry, wherein said PCI device comprises an interferometer that directs a beam to a beam splitter (37), and the method further comprises using said beam splitter (37) to direct a portion of radiation incident thereon towards a lens (37) (col. 9, lines 59-62) but does not specifically disclose the use of a microscope. Baumann et al. discloses using said microscope (2) to focus radiation from the PCI device to the eye (12) (section 0017). Therefore it would have been obvious to one having ordinary skill in the art at the time the invention was made to combine the device of Fercher with the microscope of Baumann et al. for the purpose of using the microscope as an observation system (section 0017).

Regarding claim 13, Fercher discloses, in figures 1 and 7, a method for performing biometry, wherein said PCI device further comprises a lens system, and the method further comprises using said lens (12) to focus radiation incident thereon to a portion of the eye (30) to generate a secondary radiation source on the portion of the eye (30) (col. 10, lines 5-8), wherein radiation emanating from the secondary radiation source passes through said beam splitter (37) and impinges upon said lens system (col. 12, lines 11-24) but does not specifically disclose the

Art Unit: 2873

use of a microscope. Baumann et al. discloses using said microscope (2) to focus radiation from the PCI device to the eye (12) (section 0017). Therefore it would have been obvious to one having ordinary skill in the art at the time the invention was made to combine the device of Fercher with the microscope of Baumann et al. for the purpose of using the microscope as an observation system (section 0017).

Regarding claim 14, Fercher discloses, in figures 1 and 7, a method for performing biometry, further comprising using at least one photodetector (36) to detect radiation exiting said lens system (col. 9, lines 51-54).

Regarding claim 15, Fercher discloses, in figures 1 and 7, a method for performing biometry, further comprising processing an output from said at least one photodetector (36) (col. 9, lines 51-54).

Regarding claim 16, Fercher discloses, in figures 1 and 7, a method for performing biometry, a difference between path lengths of radiation traversing arms of the interferometer equals the product of the length and refractive index of a reference eye (30) (col. 12, lines 24-30) but does not specifically disclose wherein said interferometer comprises a Michelson interferometer. It would have been obvious to modify the invention to include a Michelson interferometer, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use (In re Leshin, 125 USPQ 416). Therefore it would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the invention to include a Michelson interferometer for the purpose of its ability to allow light strikes a partially reflecting plate at an angle of 45°, the light beams reflected and transmitted by the plate are both reflected back to the plate by mirrors.

Application/Control Number: 10/724,821

Art Unit: 2873

Conclusion

Any inquiry concerning this communication or earlier communications from the

examiner should be directed to Brandi N. Thomas whose telephone number is 571-272-2341.

The examiner can normally be reached on Monday - Thursday from 6-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's

supervisor, Ricky Mack can be reached on 571-272-2333. The fax phone number for the

organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent

Application Information Retrieval (PAIR) system. Status information for published applications

may be obtained from either Private PAIR or Public PAIR. Status information for unpublished

applications is available through Private PAIR only. For more information about the PAIR

system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR

system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would

like assistance from a USPTO Customer Service Representative or access to the automated

information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

181

BNT

July 3, 2007

Brandi N Thomas Examiner Page 6

Art Unit 2873

RICKY MACK

SUPERVISORY PATENT EXAMINER